



# AURELIA® UNIQUE TPE BLUE

Powder Free TPE Examination Gloves

## AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk

+441733 615215

Supermax Healthcare Ltd,  
Unit 12-16 Titan Drive  
Fengate, Peterborough  
PE1 5XN

[aureliagloves.co.uk](http://aureliagloves.co.uk)



Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 455 1-4 Medical gloves for single use. BS EN 1186:2002 Materials and articles in contact with foodstuffs.



S	M	L	XL
48226	48227	48228	48229

### Sizes & Dimensions

### Single wall thickness

### Physical Properties

S 103+2 mm	min. 270 mm
M 105+2 mm	min. 270 mm
L 115+2 mm	min. 270 mm
XL > 120 mm	min. 270 mm



- ① min. 0.06 mm
- ② min. 0.06 mm
- ③ min. 0.05 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 400%	Min. 400%
Strength at Break	Min. 3.0N	Min. 3.0N



TPE



Powder Free



3 Years



Smooth



2.2g +/- 0.2g



Blue

